

## BIOSAFETY LEVEL 2 (BSL-2) LABORATORY PRACTICES

Outlined below are basic rules for operating a laboratory at BSL-2. This is not comprehensive. Refer to the references below for additional rules that may apply to your specific research. The Principal Investigator (PI) is ultimately responsible for the enforcement of these practices. Contact an EH&S biosafety officer for assistance at [ehsbio@uw.edu](mailto:ehsbio@uw.edu) or 206.221.7770.

<b>Training</b>	The PI and lab staff must complete <a href="#">EH&amp;S Biosafety training</a> every three years. The PI must ensure and document lab-specific biosafety training and demonstrated proficiency for lab personnel. The <a href="#">EH&amp;S Bloodborne Pathogens training</a> is required if working with human source material or BBP (e.g., Hepatitis C Virus, HIV).
<b>Access</b>	Keep lab doors closed and post a BSL-2 <a href="#">biohazard warning sign</a> while infectious agents are in use. Store infectious agents in a lockable freezer or lab room.
<b>Facilities</b>	Chairs and other furniture are covered in a non-fabric, non-porous material that can easily be decontaminated. No carpets or rugs permitted. An eyewash and a sink for hand washing must be available.
<b>Personal Protective Equipment (PPE)</b>	Wear appropriate PPE when working with infectious agents. Standard BSL-2 PPE consists of a dedicated lab coat, gloves, and eye protection. Other PPE may be required (see <a href="#">EH&amp;S PPE Assessment Guide</a> ). Do not wear PPE in public areas.
<b>Aerosol Containment</b>	Perform aerosol-generating procedures (e.g., vortexing, sonicating, pipetting, harvesting infected tissues from animals) inside a certified biological safety cabinet (BSC). Contact EH&S at <a href="mailto:ehsbio@uw.edu">ehsbio@uw.edu</a> or 206.221.7770 for evaluation.
<b>Decontamination</b>	<a href="#">Decontaminate</a> all cultures, stocks, biohazardous waste, and other potentially infectious materials prior to disposal using an appropriate method such as autoclaving or chemical disinfection. Decontaminate work surfaces and lab equipment after completion of work and after spills or splashes.
<b>Sharps</b>	Collect <a href="#">sharps</a> in a red plastic sharps container. If needles are used, do not bend, shear, break, recap, or otherwise manipulate by hand before disposal. Package non-disposable sharps in a hard-sided container for decontamination.
<b>Transport</b>	Place biohazardous waste in a leak-proof secondary container labeled with a biohazard symbol. Decontaminate the outside of the container prior to transport. PPE should not be needed for transport.
<b>References</b>	<a href="#">Biosafety in Microbiological and Biomedical Laboratories (BMBL)</a> <a href="#">UW Biosafety Manual</a>